

## DECLARATION OF CONFORMITY

**EC DECLARATION OF CONFORMITY  
 ACCORDING TO COUNCIL DIRECTIVE 89/336/EEC**

Nielsen-Kellerman Company declares that the following products, manufactured by the Nielsen-Kellerman Company at 21 Creek Circle, Boothwyn, PA 19061 USA:

<b>Product Model</b>	<b>Product Description</b>
Kestrel 2500	Pocket Weather Meter <i>Combined Wind Speed, Temperature and Barometric Pressure Measuring Instrument (Anemometer / Thermometer / Barometer)</i>
Kestrel 3500	Pocket Weather Meter <i>Combined Wind Speed, Temperature, Relative Humidity and Barometric Pressure Measuring Instrument (Anemometer / Thermometer / Hygrometer / Barometer)</i>

Are in conformity with the following standards or normative documents:

**EMISSIONS:**

EN 61000-6-3:2001 and  
 EN 61326:1997: Radiated disturbance - CISPR 22:1997,  
 Class B

**IMMUNITY:**

EN 61000-6-1:2001 and  
 EN61326:1997: Immunity, electrostatic discharge - IEC 61000-4-2:1995  
 Radiated RF interference, AM - IEC 61000-4-3:1996

Name of Authorized Person: Paul Nielsen, Vice-President, Engineering

Signature of Authorized Person: 

Test Period: 8<sup>th</sup> June 2004

Certificate Date: 18<sup>th</sup> June 2004